

TISSUE CULTURE CHAMBER - CU-30L2



*Shown with optional touchscreen controller above

Model CU-30L2 31" x 23.8" x 46.1" Exterior Dimensions (WxDxH) Temp. Range (w/ lights on) $10-44\pm0.5$ **Interior Space** 9.7 cu.ft. Total Shelving Floor Area 6.1 sq.ft. 9.1" Maximum Growing Height Up to 310 Light Intensity (6" from lamps) micromoles/m2/sec Number of Tiers 2

Applications

- *Specifically designed for tissue culture on petri plates/ dishes or in flasks
- *Many other applications exist for this product

Percival's IntellusUltra Controller

- *The IntellusUltra control system was purpose-built for controlled environments.
- *Robust, reliable, industrial-grade integrated hardware
- *Highly flexible architecture facilitates configuration, expansion and customization
- *Precise, simultaneous control of up to 7 environmental parameters
- *Industry-leading experiment protection and system diagnostics

Lighting System

- *Two tiers of lighted shelving, lit by SciWhite LEDs
- *Intensity programmable up to 310 micromoles/m2/sec of the other light irradiance measured at 6" from LEDs

Lighting System cont.

- *Programming and control of the lighting is done via IntellusUltra real time controller
- *Dimmable between 10-100% output

Airflow/Circulation

*Conditioned air circulates through a rear wall duct and is picked up by a specially designed fixed air diffuser located at the bottom of each tier (air is then delivered vertically upward at a slow speed through each shelf)

*Air diffuser insulates shelf level experiments from heat generated by underlying light fixture (this design minimizes condensation on dish lids)

Cabinet Construction

- *Interior constructed of 26-gauge galvanized steel
- *Interior floor constructed of 24-gauge polished stainless steel
- *Exterior constructed of 24-gauge galvannealed extra-smooth steel
- *NSF-complaint seam design
- *Chamber is completely self-contained, suitable for stacking one above
- *Overall wall thickness is 2"
- *Highly durable and reflective coating

Geneva Scientific LLC P.O. Box 408 Fontana, WI 53125

1-877-436-3827

Fax: 262-607-6497

Sales@Geneva-Scientific.com

Specifications are subject to change without notice.

rev. 03/22

TISSUE CULTURE CHAMBER - CU-30L2

Insulation

*Woodless construction using 2" thick foamed-in-place non-CFC Urethane insulation with 97% closed cell-structure density of 2.2 lbs/sq.ft.

Door

*One door opening of 26.8" x 29.8" providing full access to chamber interior (magnetic gasket provides a tight seal to door frame)

*Lift-off hinge design allows for simple removal of door

Interior Space

*9.7 cu.ft. with work area of 6.1 sq.ft. provided on two tiers

Shelving

*Two tiers of white epoxy coated steel wire shelving (each shelf is 16.5"D x 26.5"W)

*Shelves slide in and out easily on stainless steel rail assemblies

*Maximum growing height is 9.1"

Temperature Range

10°-44°C (±0.5°C) lights on and 2°-44°C (±0.5°C) lights off

Regulatory Standards

*Electrical Safety: UL-508A, certified and labelled by Percival Scientific under UL file number E340161

*Quality System: ISO 9001:2015, certified under DQS, Inc. under certification number 10017261

Refrigeration

*Self-contained air-cooled condensing unit with hot gas bypass system for continuous compressor operation, extended life and tight temperature control. This continuous running condensing unit ensures precise temperature control by alternating cycling refrigerant and hot gas to coil; this also prolongs life of compressor, and eliminates risk of ice build up in coil.

- *Extended stem solenoid valves for quiet and long life operation
- *Rear chamber wall mounted evaporator coils incorporates an air circulation
- *Heat rejection to ambient with standard chamber: 1,936 BTU/hr

Electrical Requirements

*Consult Geneva Scientific for electrical requirements and amperage draw.

Options

- *Additive Humidity Control with Sensor
- *Dehumidifier with Sensor
- *IntellusUltra Connect
- *Android-Based Touch Screen
- *CO2 Enrichment Package
- *Self-Contained Water-Cooled Condensing Unit
- *Dry Alarm Contacts
- *Convenience Receptacles
- *LED Lighting in Lieu of Fluorescent Lamps





Geneva Scientific LLC P.O. Box 408 Fontana, WI 53125 1-877-436-3827 Fax: 262-607-6497

Sales@Geneva-Scientific.com

Specifications are subject to change without notice.

rev. 03/22